

PIPETTOR ACCURACY CHECKS (BY VOLUME)

PERFORMED SEMI-ANNUALLY

[illegible]

Acceptable Ranges:

1 ml: 0.95 – 1.05 ml

Tolerance $\pm 5\%$

5 ml: 4.75 – 5.25 ml

Use Class A graduated cylinder

10 ml: 9.5 – 10.5 ml